


VALUE SHEET
CK-MB Control
Level 2
2.0 mL



REF CKC01.2 **LOT** DKQ44R1  27/11/2022 **VER** 1

Parameter	Unit	Target Value	Range
CK-MB	IU/L	106	84.9- 127

In order to define means and ranges, all studies were performed with Beckman AU Series. The means and limits should be adapted to each laboratory's own requirements. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

